



Approved Method: **Electrical and
Photometric
Measurements of
High-Power LEDs**

Currently in preview, click buy full version

**IES Approved Method for the
Electrical and Photometric
Measurements of High-Power LEDs**

Publication of this report
has been approved by IES.
Suggestions for revisions
should be directed to IES.

Prepared by:
**The Subcommittee on Solid-State Lighting of
the IES Testing Procedures Committee**

Copyright 2014 by the Illuminating Engineering Society of North America.

Approved by the IES Board of Directors, April 16, 2014, as a Transaction of the Illuminating Engineering Society of North America.

All rights reserved. No part of this publication may be reproduced in any form, in any electronic retrieval system or otherwise, without prior written permission of the IES.

Published by the Illuminating Engineering Society of North America, 120 Wall Street, New York, New York 10005.

IES Standards and Guides are developed through committee consensus and produced by the IES Office in New York. Careful attention is given to style and accuracy. If any errors are noted in this document, please forward them to Rita Harrold, Director of Technology, at the above address for verification and correction. The IES welcomes and urges feedback and comments.

ISBN # 978-0-87995-289-1

Printed in the United States of America.

DISCLAIMER

IES publications are developed through the consensus standards development process approved by the American National Standards Institute. This process brings together volunteers representing varied viewpoints and interests to achieve consensus on lighting recommendations. While the IES administers the process and establishes policies and procedures to promote fairness in the development of consensus, it makes no guaranty or warranty as to the accuracy or completeness of any information published herein.

The IES disclaims liability for any injury to persons or property or other damages of any nature whatsoever, whether special, indirect, consequential or compensatory, directly or indirectly resulting from the publication, use of, or reliance on this document.

In issuing and making this document available, the IES is not undertaking to render professional or other services for or on behalf of any person or entity. Nor is the IES undertaking to perform any duty owed by any person or entity to someone else. Anyone using this document should rely on his or her own independent judgment or, as appropriate, seek the advice of a competent professional in determining the exercise of reasonable care in any given circumstances.

The IES has no power, nor does it undertake, to police or enforce compliance with the contents of this document. Nor does the IES list, certify, test or inspect products, designs, or installations for compliance with this document. Any certification or statement of compliance with the requirements of this document shall not be attributable to the IES and is solely the responsibility of the certifier or maker of the statement.

**Prepared by the Subcommittee on Solid-State Lighting of
the IES Testing Procedures Committee**

LM-85 Working Group

Yoshi Ohno, *Technical Coordinator*

Dave Ellis	Jianzhong Jiao	Eric Richman
Kei Haraguchi	Rand Lee	Ralph Tuttle
Mark Hodapp	Greg McKee	Yuin Zong
Jeff Hulett	Emil Radkov	

Subcommittee on Solid State Lighting

Emil Radkov, *Chair*

John Adinolfi*	David Grandin*	Richard Li*	Ronald Rykowski*
Carl Andersen*	Michael Grather	Kurt Liepmann*	Evelyn Mahaja*
Alex Baker	Yongfeng Guan*	Joseph Liguata*	Mark Sarcoe
Peter Behnke*	Kei Haraguchi	Steven Longo	John Schutz
Robert Berger	Rudi Hechfellner*	Min-Hao Lu*	Keith Scott
Rolf Bergman	Kyle Hemmi*	Jeremy Ludyjan*	Frank Shum*
Barry Besmanoff*	Timothy Henning*	Ruiqing Ma*	Scot Solimine*
Carl Bloomfield*	Sylvia Herman*	Vikrant Mahajan	Lloyd Stafford*
Dennis Bradley*	Tanya Hernandez*	Joseph Marolla	Gregory Staples*
Eric Bretschneider	John Hickman*	Mark McCrear	Gary Steinberg
Kevin Broughton*	Yoelit Hiebert*	Greg McKee	Heidi Steward*
Jennifer Burns*	Mark Hodapp*	Jonathan Melman*	Jacki Swiernik*
Michael Buzard*	James Hospodarsky	C. Cameron Miller	David Szombattfalvy*
David Chan*	Bin Hou*	Maria Nadal*	Ted Tomonaga*
Gigi Chan	Shuming Hua*	Ronald Naus*	Ralph Tuttle
Xiaolu Chen*	Jeff Hulett	Dante Nava*	Tatsukiyo Uchida*
Jason Chesley*	Po-Chieh Hung*	David Neal*	Venkat Venkataramanan*
Jeonghyeon Choi*	Andrew Jackson	Brandon Neale*	Yaqi Wang*
Pei-Ting Chou*	Je Jang*	Paul Nie*	Joseph Welch*
Ashfaqul Chowdhury*	David Jennin	Andy Nishida*	Kurt Wilcox*
Zane Coleman*	Andrew (Sangkyoo) Jeon*	Michael O'Boyle*	Brienne Willcock*
Grace Connelly*	Jianzhong Jiao	Dan O'Hare*	Vivian Wu*
Keith Cook*	Ilina Kahn	Yoshihiro Ohno*	Wensheng Xu*
Steven Coyne*	Demetrios Karambelas	Michael O'Regan*	Jeremy Yon
James Creveling*	Mamad Kashani*	Marcel Pabst*	Richard Young*
James Dakin*	Toshihisa Kawabata*	Doosung Park*	William Young*
Ronald Daubach*	Philip Keebler*	Sagar Patel*	Gary Yu*
Lynn Davis*	Shawn Keeney*	Michael Piscitelli*	John Zhang
Marc Dyble	Tae Yeon Koo*	Michael Poplawski*	Yuin Zong*
David Egan*	Mihaly Kotrebai	Bruno Primerano*	
Phil Elizondo	Becky Kuebler	Emil Radkov	
Steve Eilersick*	Jaekwang Lee*	Sid Rane*	* Advisory Member
Dave Ellis	Rand Lee	Bipin Rao*	** Honorary Member
Caryl Kinsey Fox*	Sunghee Lee*	Irina Rasputnis*	
Calvin Galberth*	Michael Lehman*	Eric Richman	
Austin Gelder*	James Leland*	Kelvin Rong*	

IES Testing Procedures Committee

Cameron Miller, *Chair*

C. Andersen
L. Ayers*
A. Baker*
R. Berger
R. Bergin*
R. Bergman
E. Bretschneider
D. Brooks*
K. Broughton*
D. Chan*
P-T Chou*
R. Collins*
K. Curry*
R. Daubach*
D. Ellis
P. Franck*
R. Heinisch*
K. Hemmi*

T. Hernandez*
R. Horan
J. Hospodarsky
S. Hua*
P-C. Hung*
D. Husby**
S. Hutton*
A. Jackson
D. Jenkins*
J. Jiao
HS. Jung
M. Kalkas
D. Karambelas
H. Kashani*
R. Kelley*
M. Kotrebai
B. Kuebler
J. Lawton*

L. Leetzow*
K. Lerbs*
R. Levin*
I. Lewin*
R. Li*
R. Low*
J. Marella
P. McCarthy
G. McKee
C. Miller
B. Mosher
W. Newland
Y. Ohno*
G. Plank*
E. Radkov
D. Randolph
E. Richman*
M. Sapcoe

J. Schutz*
A. Smith*
D. Smith*
J. K. Son*
R. Speck**
L. Stafford*
G. Steinberg
R. Tuttle
K. Wagner*
J. Walker*
H. Wauch*
J. Welch*
K. Wilcox*
J. Yoo*
I. Zhang

* Advisory Member

** Honorary Member

Contents

Introduction	1
1.0 Scope	1
2.0 Normative References	2
3.0 Definitions	2
3.1 Device Under Test (DUT)	2
3.2 Duv	2
3.3 Goniophotometer	2
3.4 Gonio-Colorimeter	2
3.5 Gonio-Spectroradiometer	2
3.6 Heat Sink	2
3.7 High-Power LED	2
3.8 National Metrology Institute (NMI)	2
3.9 Photometer Head	2
3.10 Remote Phosphor LED Package	2
3.11 Settling Time	2
3.12 Sphere-Photometer	2
3.13 Sphere-Spectroradiometer	2
3.14 Temperature-Controlled Platform (TCP)	2
3.15 Thermal Chamber	3
4.0 Preparation for Measurements	3
4.1 Seasoning	3
4.2 Operating Orientation	3
4.3 Thermal Conditions	3
4.3.1 Junction Temperature	3
4.3.2 Temperature Control of DUT	3
4.3.3 Ambient Temperature	3
5.0 Methods of Measurement under Pulse-Mode Operation of DUT	3
5.1 General	3
5.2 Single Pulse Mode	3
5.2.1 Ambient Temperature Condition	4
5.2.2 Optical and Electrical Measurement Procedures	4
5.2.3 Estimation of Uncertainty and Correction due to Junction Temperature Error	4
5.3 Continuous Pulse Mode	5
5.3.1 Ambient Temperature Condition	5
5.3.2 Optical and Electrical Measurement Procedures	6
5.3.3 Estimation of Uncertainty and Correction due to Junction Temperature Error	6
6.0 Method of Measurement under DC-Mode Operation of DUT	6
6.1 General	6
6.2 Ambient Temperature Condition	6
6.3 Measurement Procedures	6

7.0	Optical Measurement and Equipment	7
7.1	Total Luminous Flux	7
7.1.1	Sphere-Spectroradiometer	7
7.1.2	Sphere-Photometer	8
7.1.3	Goniophotometer with a Photometer Head	8
7.1.4	Gonio-Spectroradiometer	8
7.2	Luminous Efficacy	9
7.3	Total Radiant Flux	9
7.3.1	Using Sphere-Spectroradiometer or Gonio-Spectroradiometer	9
7.3.2	Conversion Method	9
7.4	Total photon Flux	9
7.5	Luminous Intensity Distribution	9
7.6	Color Quantities	9
7.6.1	Sphere-Spectroradiometer	10
7.6.2	Gonio-Spectroradiometer or Gonio-Colorimeter	10
7.7	Wavelength Characteristics	10
7.7.1	Dominant Wavelength	10
7.7.2	Centroid Wavelength	10
7.7.3	Peak Wavelength	10
7.8	Uncertainties	10
8.0	Electrical Measurement and Equipment	10
8.1	Pulse Current Source and Pulse Current/Voltage Meter	10
8.2	Uncertainties	11
9.0	Test Report	11
	References	12
Annex A	Pulse-Mode and DC-Mode Methods for LED Operation	13
Annex B	Estimation of T_r Rise and Correlation	13
Annex C	Caution for Optical Measurement Instruments for Pulse-Mode Methods	15
Annex D	Integrating Sphere and Goniophotometer	15
Annex E	Integrating Sphere Geometries and Calibration Methods	16
Annex F	General Design of Integrating Spheres	18
Annex G	Self-Absorption Correction	19
Annex H	Calibration of Sphere-Spectroradiometer	19
Annex I	Use of Spectroradiometer	20
Annex J	Other Sources of Error	21
Annex K	Sphere Photometer System	21
Annex L	Goniophotometric Method for Total Luminous Flux	23

INTRODUCTION

This document is a guide developed for the measurement of high-power light emitting diodes (LEDs), normally in a form of LED packages, used for lighting products. High-power LEDs are those that require a heat sink for their normal operation. The light output of an LED depends strongly on its thermal conditions, in particular, the junction temperatures T_j . Junction temperature, however, is difficult to measure. Various different methods have been used to operate LEDs for photometric measurements and the results could not be compared. This document provides uniform test methods for operation of high-power LEDs and test methods for photometric and colorimetric measurement of high-power LEDs.

The photometric measurement of high-power LEDs has been difficult because they are highly sensitive to thermal operating conditions, and there has been a lack of common methods that can be used by both LED manufacturers and users to acquire reproducible results. LED manufacturers normally use pulse operation, whereby, LEDs are measured with no heat sink and with the underlying assumption that the junction temperature under these conditions is equal to the room temperature, typically 25°C. Therefore, published LED specifications are normally at junction temperature of 25°C. High-power LEDs in actual lighting products, however, are operating in DC and at much higher temperatures (typically junction temperature is 60°C to over 100°C), where their photometric and colorimetric values tend to deviate significantly from the room temperature condition. To assist users, LED manufacturers make efforts to provide thermal characteristics data at higher operating temperatures; however, because LEDs are usually binned by LED manufacturers for their optical and electrical characteristics at T_j equal to 25°C, manufacturer data for higher operating temperatures is of limited use. There have been no standard methods for measuring high-power LEDs at high temperatures. This document provides reproducible measurement methods of LEDs at a given junction temperature in pulse or DC mode and provides the grounds for specification of LEDs at high temperature conditions.

Lighting product manufacturers often need to know the performance of LEDs operating in full rated DC current at a thermal equilibrium at much higher temperatures than 25°C. To set or measure thermal conditions of the LED, “case temperature”, “pin temperature”, “board temperature”, “solder-point temperature”, or “heat sink temperature” are commonly used depending on the type of LED. While these methods are useful to reproduce the same condition

for the particular LED, the results using these different methods cannot be compared with each other and cannot be reconciled into a universal standard method. Due to the optical characteristics of LEDs and their dependence on junction temperature, the only way to obtain reproducible results universally for all types of LEDs is by setting them to a specified junction temperature. The test method described in this document is to set the LED under test to a pre-determined junction temperature, for measurement at either pulse mode or DC mode operation. Such a method can establish equivalence of results between the pulse mode tests (normally performed by LED manufacturers) and the DC mode tests (often preferred by users of LEDs).

The photometric and radiometric information typically required for high-power LEDs for lighting products is total luminous flux (lumens), total radiant flux (watts), total photon flux (moles), and luminous efficacy (lm/W). The colorimetric information includes chromaticity coordinates (for all LEDs); correlated color temperature (CCT) Duv, and color rendering index (CRI) for white LEDs; dominant wavelength, centroid wavelength, and peak wavelengths (for colored LEDs). For the purpose of this document, the determination of these values are referred to as optical measurements.

The electrical characteristics typically required for high-power LEDs for lighting products are, input DC current, forward voltage, and input power. For the purpose of this document, the determination of these values are referred to as electrical measurements.

For special purposes, it may be useful to determine the characteristics of LEDs when they are operated at conditions other than the nominal conditions described in this approved method. When measurements are conducted at conditions other than the nominal conditions, the results are valid only for the particular conditions under which they were obtained; these conditions shall be stated in the test report.

1.0 SCOPE

This document describes the procedures to be followed and precautions to be observed in performing accurate measurements of total luminous flux, total radiant flux (optical power), total photon flux, electrical power, luminous efficacy, color quantities, and wavelength characteristics of high-power light emitting diodes (LEDs) including white LEDs as well as single color LEDs. This document covers LED packages (defined in ANSI/IES RP-16-10) including those with multiple chips and remote-phosphor LED pack-

ages. This document covers measurement under pulse operation as well as steady DC operation of LEDs, and in all cases, the thermal condition of LEDs refers to their junction temperature. The approved methods apply to laboratory measurements.

This document does not cover LED arrays or modules, nor LED lighting products; it does not cover AC driven LEDs, and does not apply to measurements in LED manufacturer's production control nor relative measurements of LED thermal characteristics.

2.0 NORMATIVE REFERENCES

2.1 ANSI/IES RP-16-10, *Nomenclature and Definitions for Illuminating Engineering.*

2.2 CIE S017/E:2011 ILV, *International Lighting Vocabulary.*

2.3 ISO 23539:2005(E)/CIE S 010/E 2004, *Photometry - The CIE System of Physical Photometry.*

2.4 ISO 11664-1:2007(E)/CIE S 014-1/E 2007, *Colorimetry – Part 1: CIE Standard Colorimetric Observers*

3.0 DEFINITIONS

3.1 Device Under Test (DUT)

High-power LED package under test.

3.2 Duv

The closest distance from the chromaticity coordinate of the light source to the Planckian locus on the CIE (u' , $2/3 v'$) coordinates with "+" sign for above and "-" sign for below the Planckian locus.

3.3 Goniophotometer

Photometer for measuring the directional light distribution characteristics of sources, luminaires, media, or surfaces. In this document, "goniophotometer" includes goni-colorimeter and goni-spectroradiometer.

3.4 Goni-Colorimeter

Goniophotometer equipped with a tristimulus colorimeter head as the detector.

3.5 Goni-Spectroradiometer

Goniophotometer equipped with a spectroradiometer as the detector.

3.6 Heat Sink

A device attached to an LED assembly (package, array or module) to dissipate heat.

3.7 High-Power LED

LED package that requires a heat sink or other means of thermal management for its normal operation.

3.8 National Metrology Institute (NMI)

A national laboratory that maintains the SI units for the country and authorized to disseminate calibration standards for measurements.

3.9 Photometer Head

A unit containing a detector, a $V(\lambda)$ -correction filter, and any additional components (aperture, diffuser, amplifier, etc.) within the unit.

3.10 Remote Phosphor LED Package

An LED package containing a phosphor that is spatially separated from the pump or source radiation element, but contained within the LED package.

3.11 Settling Time

Time needed for a signal to reach a stable level after a fast transition.

3.12 Sphere-Photometer

Integrating sphere equipped with a photometer head as the detector.

3.13 Sphere-Spectroradiometer

Integrating sphere equipped with a spectroradiometer as the detector.

Note: This type of instrument measures total spectral radiant flux (unit: W/nm), from which total luminous flux, total radiant flux, and color quantities are obtained.

3.14 Temperature-Controlled Platform (TCP)

A device that incorporates a cooling and/or heating component(s) to maintain its temperature at a set value. TCP is used to control temperature of a high-