



**LIGHTING SCIENCE:
MEASUREMENT OF LIGHT –
THE SCIENCE OF PHOTOMETRY**
AN AMERICAN NATIONAL STANDARD

Currently in preview, click buy full version



ANSI/IES LS-4-20

**LIGHTING SCIENCE:
MEASUREMENT OF LIGHT –
THE SCIENCE OF PHOTOMETRY
AN AMERICAN NATIONAL STANDARD**

Publication of this Committee report
has been approved by IES.
Suggestions for revisions
should be directed to IES.

**Prepared by the
IES Testing Procedures Committee**



Copyright 2020 by the Illuminating Engineering Society.

Approved by the IES Standards Committee January 6, 2020 as a Transaction of the Illuminating Engineering Society.

Approved March 13, 2020 as an American National Standard.

All rights reserved. No part of this publication may be reproduced in any form, in any electronic retrieval system or otherwise, without prior written permission of the IES.

Published by the Illuminating Engineering Society, 120 Wall Street, New York, New York 10005

IES Standards are developed through committee consensus and produced by the IES Office in New York. Careful attention is given to style and accuracy. If any errors are noted in this document, they should be forwarded to Brian Liebel, Director Standards, at standards@ies.org or the above address for verification and correction. The IES welcomes and urges feedback and comments.

Printed in the United States of America.

ISBN# 978-0-87995-289-1

DISCLAIMER

IES publications are developed through the consensus standards development process approved by the American National Standards Institute. This process brings together volunteers representing varied viewpoints and interests to achieve consensus on lighting recommendations. While the IES administers the process and establishes policies and procedures to promote fairness in the development of consensus, it makes no guaranty or warranty as to the accuracy or completeness of any information published herein.

The IES disclaims liability for any injury to persons or property or other damages of any nature whatsoever, whether special, indirect, consequential or compensatory, directly or indirectly resulting from the publication, use of, or reliance on this document.

In issuing and making this document available, the IES is not undertaking to render professional or other services for or on behalf of any person or entity. Nor is the IES undertaking to perform any duty owed by any person or entity to someone else. Anyone using this document should rely on his or her own independent judgment or, as appropriate, seek the advice of a competent professional in determining the exercise of reasonable care in any given circumstances.

The IES has no power, nor does it undertake, to police or enforce compliance with the contents of this document. Nor does the IES list, certify, test or inspect products, designs, or installations for compliance with this document. Any certification or statement of compliance with the requirements of this document shall not be attributable to the IES and is solely the responsibility of the certifier or maker of the statement.

AMERICAN NATIONAL STANDARD

Approval of an American National Standard requires verification by ANSI that the requirements for due process, consensus, and other criteria have been met by the standards developer.

Consensus is established when, in the judgment of the ANSI Board of Standards Review, substantial agreement has been reached by directly and materially affected interests. Substantial agreement means much more than a simple majority, but not necessarily unanimity. Consensus requires that all views and objections be considered, and that a concerted effort be made toward their resolution.

The use of American National Standards is completely voluntary; their existence does not in any respect preclude anyone, whether that person has approved the standards or not, from manufacturing, marketing, purchasing, or using products, processes, or procedures not conforming to the standards.

The American National Standards Institute does not develop standards and will in no circumstances give an interpretation to any American National Standard. Moreover, no person shall have the right or authority to issue an interpretation of an American National Standard in the name of the American National Standards Institute. Requests for interpretations should be addressed to the secretariat or sponsor whose name appears on the title page of this standard.

CAUTION NOTICE: This American National Standard may be revised at any time. The procedures of the American National Standards Institute require that action be taken to reaffirm, revise, or withdraw this standard no later than five years from the date of approval. Purchasers of American National Standards may receive current information on all standards by calling or writing the American National Standards Institute.

Prepared by the IES Testing Procedures Committee.

Becky Kuebler, *Chair*

Andrew Jackson, *Vice Chair*

David N. Randolph, *Secretary*

Jianzhong Jiao, *Treasurer*

Members

C. K. Andersen

R. P. Bergin

R. S. Bergman

E. Bretschneider

P. Elizondo

K. C. Fletcher

M. L. Grather

Y. Hiebert

J. Hospodarsky

J. N. Hulett

P.-C. Hung

M. Kotrebai

J. E. Leland

S. Longo

J. P. Marella

P. McCarthy

G. McKee

C. C. Miller

E. Radkov

M. B. Saborio

Advisory Members

L. M. Ayers

J. Baker

C. A. Bloomfield

P.-T. Chou

P. Cruz

M. Damle

L. Davis

J. J. Demirjian

M. E. Duffy

V. Eberhard

A. A. Feldman

J. Frazer

K. J. Hemmi

S. Hua

G. John

J. Juhasz

H. Kashaninejad

T. Kawabata

J. D. Kramer

K. C. Lerbs

K. M. Liepmann

J. Lockner

Y. Ohno

E. Page

D. Park

E. S. Perkins

M. Piscitelli

T. W. Rasinski

D. Rogers

M. P. Rorer

T. Schneider

A. W. Sarras

S. A. Steinberg

L. Swainston

J. S. Swiernik

A. Thorseth

R. C. Tuttle

J. C. Vollers

J. E. Walker

Y. Zong

CONTENTS

1.0 Introduction and Scope	1
1.1 Introduction	1
1.2 Scope	1
2.0 Traceability, Accreditation, and Photometric Standards	2
2.1 Traceability and Accreditation	2
2.2 Photometric Standards	3
2.2.1 Types of Standards	3
3.0 Visual Photometry	4
4.0 Physical Photometry	4
4.1 Types of Detectors	4
4.1.1 Phototubes	4
4.1.2 Solid-State Detectors	5
4.2 Detector Spectral Response	5
4.3 Environmental Factors	6
4.3.1 Temperature Effects	6
4.3.2 Transient Effects	6
4.3.3 Effect of Pulsed or Cyclical Variation of Light	6
4.3.4 Magnetic Fields	6
4.3.5 Electrical Interference	6
5.0 Absolute, Relative, and Substitution Photometry	6
5.1 Absolute Photometry	6
5.2 Relative Photometry	7
5.3 Substitution Photometry	7
6.0 Instrument Accuracy	7
6.1 Factors for All Instruments	7
6.1.1 Spectral Correction Error, f_1'	7
6.1.2 Linearity Error, f_3	8
6.1.3 Display Error, f_4	8
6.1.4 Fatigue Error, f_5	8
6.2 Illuminance Meter Cosine Response Error, f_2	8
6.3 Luminance Meter Surround Field Error, $f_2(u)$	9

7.0	Measuring Spectra	9
7.1	Using Spectroradiometers	10
7.1.1	Measurement of Spectral Power Distribution (SPD)	10
7.1.2	Spectral Reflectance and Transmittance	11
8.0	Measuring Illuminance	11
8.1	Illuminance Meters	11
8.2	Angular Response	11
8.3	Spectral Response	12
9.0	Measuring Luminous Intensity	12
9.1	Optical Bench Photometry	12
9.2	Distribution Photometry	12
10.0	Measuring Flux	13
10.1	Integrating Sphere	14
11.0	Measuring Luminance	15
11.1	Spot Luminance Meters	15
11.1.1	Beam Splitter Spot Meter	15
11.1.2	Aperture Mirror Photometers	16
11.2	Digital Luminance Meters	16
12.0	Measuring Reflectance and Transmittance	17
12.1	Reflectometers (and Transmittometers)	17
12.2	Field Measurement of Reflectance	18
13.0	Lamp Photometry	18
13.1	Characterizing Lamps	19
13.2	Lamp Testing	19
14.0	Luminaire Photometry	20
14.1	Far-Field and Near-Field Photometry	20
14.2	Absolute Luminaire Photometry	20
14.3	Relative Luminaire Photometry	20
14.4	Characterizing Luminaires	20
14.5	Luminaire Photometric Testing	21

14.6	Derived Photometric Characteristics	21
14.6.1	Zonal Lumens	21
14.6.2	Luminaire Efficiency and Photometric Efficiency	21
14.6.3	Average Luminance	22
14.6.4	Beam Type and Characterization	22
14.6.5	Coefficients of Utilization	22
15.0	Field Measurements	23
15.1	Interior Measurements	23
15.1.1	Average Illuminance	23
15.1.2	Illuminance at a Point	25
15.1.3	Luminance	25
15.2	Outdoor Measurements	25
	References	27

Currently in preview, click buy full versi

When you can measure what you are speaking about, and express it in numbers, you know something about it.

– Lord Kelvin 1883, engineer and mathematical physicist

1.0 Introduction and Scope

1.1 Introduction

Lighting is anchored to meaningful visual phenomena by way of the definition of *light* it adopts: the joining of radiant power, a physical quantity, with visual response, a psychophysical quantity. The utility of the definition allows light to be measured and calculated; that is, light has the aspects of *quantity* that permit the engineering activities of measurement and prediction by calculation. This analytic aspect of light and lighting allows successful experience to be recorded and become quantity recommendations for other lighting projects, allows lighting equipment to be characterized in ways useful to designers, and allows predictions of likely outcomes of proposed lighting designs.

Optical radiation generally refers to all radiation that can be measured using certain techniques and equipment (mirrors, lenses, filters, diffraction gratings, prisms). Thus, visible, ultraviolet (UV), and infrared (IR) radiation are collectively considered as optical radiation.

The measurement of optical radiation, called *radiometry*, is the science of measuring radiant quantities and is part of the general science of measurement, metrology.

Photometry, a special branch of radiometry, is the measurement of radiation accounting for human visual response. The International Commission on Illumination (CIE) defines a “standard observer” in part by the photopic luminous efficiency function of wavelength, $V(\lambda)$. The definition of this standard observer quantifies the visual response and defines the spectral response that photometric measurement equipment needs to exhibit.¹

This standard observer response curve is used as a weighting function applied to the spectral power distribution (SPD) of the optical radiation being

Sidebar: A Historical Note

Photometry is a word first used by Johann Heinrich Lambert as the title to his 1760 Latin treatise on the measurement of light. He coined it by combining the Greek words for *light* ($\phi\omega\varsigma$) and *measure* ($\mu\epsilon\tau\omicron\nu$). Lambert’s word soon found its way into European languages.

measured. The summation across all wavelengths of the weighted SPD defines luminous flux.¹ The weighting and summation are the very core of photometry. Though it is globally accepted and used, $V(\lambda)$ is a compromise that always assumes the same predictable correlation of physical measurements with visual response. However, there are circumstances where photometric quantities are poor predictors of visual response. (Refer to ANSI/IES E-8-20, *Lighting Science: Vision – Perceptions and Performance*.) Thus, a basic understanding of photometry is essential to the balance that needs to be struck by a lighting designer between measurement on one hand and visual experience on the other.

Photometry and radiometry are used to determine properties of lighting equipment and materials, as well as aspects of the performance of lighting systems. Some of these measurements require photometric standards (for either sources or detectors) and are usually performed in a photometric laboratory; some are accomplished with equipment designed for field use. **Table 1-1** shows the most common types of photometric and radiometric measurement, along with the equipment and usual place of measurement.

1.2 Scope

This Lighting Science (LS) document describes the various types of photometry and photometric instrumentation, including laboratory and field equipment and measurement types, and instructions for some types of field measurements. However, it does not provide instructions or methodology for performing laboratory tests. For that kind of information, the reader is referred to the IES Lighting Measurement (LM) series of documents.