

# JIS

JAPANESE  
INDUSTRIAL  
STANDARD

Translated and Published by  
Japanese Standards Association

---

---

**JIS G 2404** : 2015

(JAA/JSA)

**Methods for chemical analysis of  
aluminium dross for iron and steel  
making**

---

ICS 77.040.01;77.120.10

Reference number : **JIS G 2404 : 2015 (E)**

Date of Establishment: 2015-03-24

Date of Public Notice in Official Gazette: 2015-09-24

Investigated by: Japanese Industrial Standards Committee

Standards Board for ISO area

Technical Committee on Metal and Inorganic

Materials

---

JIS G 2404:2015, First English edition published in 2016-03

Translated and published by: Japanese Standards Association  
Mita MT Building, 3-13-12, Mita, Minato-ku, Tokyo, 108-0073 JAPAN

---

In the event of any doubts arising as to the contents,  
the original JIS is to be the final authority.

© JSA 2016

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

Printed in Japan

NH/AT

## Contents

		Page
1	Scope .....	1
2	Normative references .....	1
3	General requirements .....	1
4	Determination method of metallic aluminium .....	2
4.1	Division of determination methods .....	2
4.2	ICP atomic emission spectrometry for decomposed and separated metallic aluminium .....	2
4.3	Zinc back-titration method for decomposed and separated metallic aluminium .....	5
4.4	Volumetric method by gas dissolved in hydrochloric acid .....	7
4.5	Measurement method for reaction solution temperature .....	12
5	Determination method of nitrogen .....	14
5.1	Determination method .....	14
5.2	Summary .....	14
5.3	Reagents .....	14
5.4	Apparatus for steam distillation .....	16
5.5	Sample for chemical analysis .....	16
5.6	Amount of sample weighed out .....	16
5.7	Operation .....	16
5.8	Blank test .....	18
5.9	Calculation .....	18
6	Determination method of chloride ion .....	18
6.1	Division of determination methods .....	18
6.2	Ion chromatography .....	19
6.3	Silver nitrate titration method .....	22
7	Determination method of total silicon .....	26
7.1	Division of determination methods .....	26
7.2	Molybdo silicic acid absorption spectrophotometry .....	26
7.3	Silicon dioxide gravimetric method .....	29
8	Determination method of aluminium oxide .....	31
8.1	Division of determination methods .....	31
8.2	ICP atomic emission spectrometry for decomposed and separated metallic aluminium .....	32
8.3	Zinc back-titration method for decomposed and separated metallic aluminium .....	35
9	Determination method of carbon .....	37

9.1	Determination method .....	37
9.2	Summary .....	37
9.3	Reagents .....	37
9.4	Implements and materials .....	39
9.5	Apparatus .....	39
9.6	Sample for analysis .....	41
9.7	Amount of sample weighed out .....	41
9.8	Operation .....	41
9.9	Blank test .....	44
9.10	Preparation of calibration curve .....	44
9.11	Correction of calibration curve .....	48
9.12	Calculation .....	49
10	Determination method of moisture .....	49
10.1	Determination method .....	49
10.2	Summary .....	49
10.3	Sample .....	49
10.4	Apparatus and instruments .....	49
10.5	Operation .....	50
10.6	Calculation .....	50
10.7	Tolerance .....	50
10.8	Analysed value .....	50
11	Precautions on safety and hygiene, and environment .....	50

## Foreword

This translation has been made based on the original Japanese Industrial Standard established by the Minister of Economy, Trade and Industry through deliberations at the Japanese Industrial Standards Committee according to the proposal for establishment of Japanese Industrial Standard submitted by Japan Aluminium Association (JAA)/Japanese Standards Association (JSA) with the draft being attached, based on the provision of Article 12 Clause 1 of the Industrial Standardization Law.

Consequently, this Standard partially replaces **JIS G 2402**:2009 and **JIS G 2403**:2003.

This **JIS** document is protected by the Copyright Law.

Attention is drawn to the possibility that some parts of this Standard may conflict with patent rights, applications for a patent after opening to the public or utility model rights. The relevant Minister and the Japanese Industrial Standards Committee are not responsible for identifying any of such patent rights, applications for a patent after opening to the public or utility model rights.

# Methods for chemical analysis of aluminium dross for iron and steel making

## 1 Scope

This Japanese Industrial Standard specifies the determination methods for the metallic aluminium, nitrogen, chloride ion, total silicon, aluminium oxide and carbon, and moisture in aluminium dross for iron and steel making used as the flux at the time of making iron and steel.

## 2 Normative references

The following standards contain provisions which, through reference in this text, constitute provisions of this Standard. The most recent editions of the standards (including amendments) indicated below shall be applied.

JIS G 2403 *Aluminium dross for iron and steel making – Sampling and sample preparation procedures*

JIS K 0050 *General rules for chemical analysis*

JIS K 0115 *General rules for molecular absorption spectrometry*

JIS K 0116 *General rules for atomic emission spectrometry*

JIS K 0127 *General rules for ion chromatography*

JIS K 0970 *Piston pipettes*

JIS K 1101 *Oxygen*

JIS K 8005 *Reference materials for volumetric analysis*

JIS R 1306 *Porcelain combustion boats for chemical analysis*

JIS Z 2615 *General rules for determination of carbon in metallic materials*

## 3 General requirements

General requirements common to the chemical analysis shall be in accordance with **JIS K 0050** and the following apply.

- a) The sampling and the sample preparation shall be in accordance with **JIS G 2403**.
- b) Water to be used shall be A2 or A3 in grade and quality specified in Annex D of **JIS K 0050**.
- c) When the corresponding **JIS** exists, reagents and gases of the highest grade or those appropriate for the purpose shall be used. When the corresponding **JIS** does not exist, those do not interfere with analysis in quality shall be used.
- d) General requirements common to ICP atomic emission spectrometry shall be in accordance with **JIS K 0116**.