

JIS

JAPANESE
INDUSTRIAL
STANDARD

Translated and Published by
Japanese Standards Association

JIS K 0117 : 2017

(JAIMA/JSA)

**General rules for infrared
spectrophotometric analysis**

ICS 71.040.50

Reference number : **JIS K 0117 : 2017 (E)**

Date of Establishment: 1966-11-01

Date of Revision: 2017-03-21

Date of Public Notice in Official Gazette: 2017-03-21

Investigated by: Japanese Industrial Standards Committee
Standards Board for ISO area

Technical Committee on Chemical Products and
Analytical Methods

JIS K 0117:2017, First English edition published in 2017-12

Translated and published by: Japanese Standards Association
Mita MT Building, 3-13-12, Mita, Minato-ku, Tokyo, 108-0073 JAPAN

In the event of any doubts arising as to the contents,
the original JIS is to be the final authority.

© JSA 2017

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

Printed in Japan

NH/AT

Contents

| | | Page |
|------|--|------|
| 1 | Scope | 1 |
| 2 | Normative references | 1 |
| 3 | Terms and definitions | 1 |
| 4 | Apparatus | 2 |
| 4.1 | General | 2 |
| 4.2 | Infrared spectrophotometer | 4 |
| 4.3 | Accessories | 9 |
| 4.4 | Additional functions | 17 |
| 5 | Sample preparation | 21 |
| 5.1 | Notes | 21 |
| 5.2 | Solid samples | 21 |
| 5.3 | Powder samples | 23 |
| 5.4 | Liquid samples | 23 |
| 5.5 | Gaseous samples | 24 |
| 6 | Operation procedures | 24 |
| 6.1 | Installation of apparatus | 24 |
| 6.2 | Measurement procedures | 25 |
| 6.3 | Correction and inspection methods of spectrophotometer | 26 |
| 7 | Qualitative analysis | 27 |
| 7.1 | Method based on analysis of absorption spectrum | 28 |
| 7.2 | Method based on comparison with the spectrum of known compound | 28 |
| 8 | Quantitative analysis | 29 |
| 8.1 | Determination method | 29 |
| 8.2 | Expression of determination values | 31 |
| 9 | Guarantee of quality of data | 31 |
| 9.1 | General | 31 |
| 9.2 | Checks on apparatus | 31 |
| 9.3 | Performance validation of apparatus | 31 |
| 9.4 | Optimum method of sample preparation | 32 |
| 9.5 | Quality control of data obtained in quantitative analysis | 32 |
| 10 | Safety and maintenance | 33 |
| 10.1 | Safety | 33 |
| 10.2 | Maintenance | 34 |
| 11 | Record of measurement details | 34 |

| | | |
|------|--|----|
| 12 | Matters to be included in individual standards | 34 |
| 12.1 | Qualitative analysis | 34 |
| 12.2 | Quantitative analysis..... | 35 |

Currently in preview, click buy full version

Foreword

This translation has been made based on the original Japanese Industrial Standard revised by the Minister of Economy, Trade and Industry through deliberations at the Japanese Industrial Standards Committee as the result of proposal for revision of Japanese Industrial Standard submitted by Japan Analytical Instruments Manufacturers' Association (JAIMA)/Japanese Standards Association (JSA) with the draft being attached, based on the provision of Article 12 Clause 1 of the Industrial Standardization Law applicable to the case of revision by the provision of Article 14.

Consequently **JIS K 0117:2000** is replaced with this Standard.

This **JIS** document is protected by the Copyright Law.

Attention is drawn to the possibility that some parts of this Standard may conflict with patent rights, applications for a patent after opening to the public or utility model rights. The relevant Minister and the Japanese Industrial Standards Committee are not responsible for identifying any of such patent rights, applications for a patent after opening to the public or utility model rights.

General rules for infrared spectrophotometric analysis

1 Scope

This Japanese Industrial Standard specifies the general rules applicable to the qualitative analysis and quantitative analysis of inorganic and organic substances using infrared spectrophotometer¹⁾.

Note¹⁾ In a broad definition, the term “infrared rays” means electromagnetic waves having wavelength between visible ray and microwaves. In this Standard, however, it is treated as rays in the wavenumber range of 4 000 cm⁻¹ to 400 cm⁻¹ (wavelengths 2.5 μm to 25 μm).

2 Normative references

The following standards contain provisions which, through reference in this text, constitute provisions of this Standard. The most recent editions of the standards (including amendments) indicated below shall be applied.

JIS K 0050 *General rules for chemical analysis*

JIS K 0211 *Technical terms for analytical chemistry (General part)*

JIS K 0212 *Technical terms for analytical chemistry (optical part)*

JIS K 0215 *Technical terms for analytical chemistry (Analytical instrument part)*

3 Terms and definitions

For the purposes of this Standard, the terms and definitions given in **JIS K 0211**, **JIS K 0212** and **JIS K 0215**, and the following apply.

3.1 interferogram

diagram expressing the signal sent from Michelson interferometer, obtained by plotting the optical path difference of light on abscissa and the light intensity on ordinate

3.2 apodization

mathematical operation to overlap the adequate function on interferogram, in order to reduce the spectrum distortion caused by the finite scanning distance of interferometer

3.3 absorbance

numerical value²⁾ expressed by the common logarithm of the ratio between the intensity (I_0) of the incident light directed at a sample and the intensity (I) of the light transmitted through it

Note²⁾ Absorbance (Abs) is expressed as follows.

$$\text{Absorbance (Abs)} = -\log_{10}(I/I_0)$$