

JIS

JAPANESE
INDUSTRIAL
STANDARD

Translated and Published by
Japanese Standards Association

JIS K 0311 : 2020

Method for determination of tetra-through octachlorodibenzo-*a*-dioxins, tetra-through octachlorodibenzofurans and dioxin-like polychlorinated biphenyls in stationary source emissions

This document has incorporated the erratum to the Japanese version issued in May, 2021.

ICS 13.040.40;71.040.40

Reference number : JIS K 0311 : 2020 (E)

K 0311 : 2020

Date of Establishment: 1999-09-20

Date of Revision: 2020-03-23

Date of Public Notice in Official Gazette: 2020-03-23

Investigated by: Japanese Industrial Standards Committee
Standards Board for ISO area
Technical Committee on Chemical Products and
Analytical Methods

JIS K 0311:2020, First English edition published in 2021-07

Translated and published by: Japanese Standards Association
Mita MT Building, 3-13-12, Mita, Minato-ku, Tokyo, 108-0073 JAPAN

In the event of any doubts arising as to the contents,
the original JIS is to be the final authority.

© JSA 2021

All rights reserved. Unless otherwise specified, no part of this publication may be reproduced or utilized in any form or by any means, electronic or mechanical, including photocopying and microfilm, without permission in writing from the publisher.

Printed in Japan

KK/AT

PROTECTED BY COPYRIGHT

Contents

	Page
Introduction	1
1 Scope	1
2 Normative references	1
3 Terms, definitions and abbreviations	2
3.1 Terms and definitions	2
3.2 Abbreviations	4
4 Outline of determination method	5
5 Collection of sample gas	6
5.1 Outline of collection of sample gas	6
5.2 Gas sampling apparatus	6
5.3 Reagents	8
5.4 Preparations for the collection of sample gas	11
5.5 Amount of sample gas collected	12
5.6 Sampling operations	12
5.7 Recovery and storage of the sample	13
5.8 Calculation of amount of sample gas collected	14
5.9 Recording of collection of sample gas	14
6 Pretreatment of the sample	14
6.1 Outline of the pretreatment of the sample	14
6.2 Reagents and materials	16
6.3 Instrument and apparatus	18
6.4 Pretreatment operations	19
7 Identification and determination	28
7.1 Outline	28
7.2 Reagents and apparatus	28
7.3 Measurement operations	31
7.4 Identification and determination of dioxins	36
7.5 Limit of detection and minimum limit of determination	39
7.6 Confirmation of the recovery	42
8 Report of the results	43
8.1 Methods of expression for analytical results	43
8.2 Unit of concentration	44
8.3 Conversion to toxic equivalents	44
8.4 Handling of numerical values	45
9 Quality control of the measured data	45

9.1	General	45
9.2	Assurance of reliability of measured data	45
9.3	Requirements for measurement operations	48
9.4	Record of measurement operations	52
9.5	Report on the precision control	52
Annex A (normative)	Gas sampling apparatus	54
Annex B (informative)	Example of use of internal standard	60
Annex C (informative)	Examples of measurement conditions of GC-MS and chromatogram	66
Annex D (informative)	Correction by oxygen concentration	77
Annex E (informative)	Conversion to TEQ	80

Foreword

This Japanese Industrial Standard has been revised by the Minister of Economy, Trade and Industry, through deliberations at the Japanese Industrial Standards Committee in accordance with the Industrial Standardization Act. This edition replaces the previous edition (**JIS K 0311**:2008), which has been technically revised.

This **JIS** document is protected by the Copyright Act.

Attention is drawn to the possibility that some parts of this Standard may conflict with patent rights, published patent application or utility model rights. The relevant Minister and the Japanese Industrial Standards Committee are not responsible for identifying any of such patent rights, published patent application or utility model rights.

Method for determination of tetra-through octachlorodibenzo-*p*-dioxins, tetra-through octachlorodibenzofurans and dioxin-like polychlorinated biphenyls in stationary source emissions

Introduction

This Japanese Industrial Standard was established in 1999, and has gone through three revisions including this one. After the previous revision was made in 2008, organic solvents used for extraction have been designated as specified chemical substances in the Industrial Safety and Health Act. In step with this, this Standard has been revised in order to ensure the safety of analytical workers.

No corresponding International Standard has been established at this point.

1 Scope

This Standard specifies a method for the determination of tetra-through octachlorodibenzo-*p*-dioxins, tetra-through octachlorodibenzofurans and dioxin-like PCBs in flue gas which are generated by combustion and chemical reactions, and discharged to a flue, stack or duct (hereafter referred to as duct) in stationary source emissions using a gas chromatograph/mass spectrometer (hereafter referred to as GC-MS). The GC-MS employed in this Standard shall be the double focusing mass spectrometer (hereafter referred to as MS) where the capillary column of a gas chromatograph (hereafter referred to as GC) is used and its resolution is 10 000 or more.

WARNING Since dioxins are very toxic, inhalation, accidental ingestion, direct contact to skin, etc. shall be avoided as much as possible, and ventilation of pretreatment room and analysis room and control of waste liquid and waste shall be sufficiently carried out. Since the health of the measurer may be damaged by inhalation or accidental ingestion of other drugs, solvents, etc., the handling shall be carried out as carefully as possible, and attention shall be paid to sufficient ventilation in the laboratory.

2 Normative references

The following standards contain provisions which, through reference in this text, constitute provisions of this Standard. The most recent editions of the standards (including amendments) indicated below shall be applied.

JIS K 0095 *Methods for sampling of flue gas*

JIS K 0114 *General rules for gas chromatography*

JIS K 0123 *General rules for gas chromatography/mass spectrometry*

JIS K 0211 *Technical terms for analytical chemistry (General part)*

JIS K 0215 *Technical terms for analytical chemistry (Analytical instrument part)*