

NEMA LC 1-2007 (R2013, 2018)

---

# Test Procedure for Compatibility of Hearing Aids and Ultrasonic Lighting Control Devices



**NEMA Standards Publication LC 1-2007 (R2012, 2018)**

*Test Procedure for Compatibility of  
Hearing Aids and Ultrasonic Lighting Control Devices*

*Published by:*

**National Electrical Manufacturers Association**

1300 North 17<sup>th</sup> Street, Suite 900  
Rosslyn, VA 22209

[www.nema.org](http://www.nema.org)

© 2018 National Electrical Manufacturers Association. All rights, including translation into other languages, reserved under the Universal Copyright Convention, the Berne Convention for the Protection of Literary and Artistic Works, and the International and Pan American copyright conventions.

## NOTICE AND DISCLAIMER

The information in this publication was considered technically sound by a consensus among persons engaged in its development at the time it was approved. Consensus does not necessarily mean there was unanimous agreement among every person participating in the development process.

The National Electrical Manufacturers Association (NEMA) Standards and guideline publications, of which the document herein is one, are developed through a voluntary Standards development process. This process brings together volunteers and/or seeks out the views of persons who have an interest in the topic covered by this publication. Although NEMA administers the process and establishes rules to promote fairness in the development of consensus, it does not write the documents, nor does it independently test, evaluate, or verify the accuracy or completeness of any information or the soundness of any judgments contained in its Standards and guideline publications.

NEMA disclaims liability for any personal injury, property, or other damages of any nature, whether special, indirect, consequential, or compensatory, directly or indirectly resulting from the publication, use of, application, or reliance on this document. NEMA disclaims and makes no warranty or warranty, express or implied, as to the accuracy or completeness of any information published herein, and disclaims and makes no warranty that the information in this document will fulfill any particular purpose(s) or need(s). NEMA does not undertake to guarantee the performance of any individual manufacturer's or seller's products or services by virtue of this Standard or guide.

In publishing and making this document available, NEMA is not undertaking to render professional or other services for or on behalf of any person or entity, nor is NEMA undertaking to perform any duty owed by any person or entity to someone else. Anyone using this document should rely on his or her own independent judgment or, as appropriate, seek the advice of a competent professional in determining the exercise of reasonable care in any given circumstance. Information and other Standards on the topic covered by this publication may be available from other sources, which the user may wish to consult for additional views or information not covered by this publication.

NEMA has no power, nor does it undertake to police or enforce compliance with the contents of this document. NEMA does not certify, test, or inspect products, designs, or installations for safety or health purposes. Any certification or other statement of compliance with any health- or safety-related information in this document shall not be attributable to NEMA and is solely the responsibility of the certifier or maker of the statement.

**CONTENTS**

Foreword .....	iv
<b>Section 1 General</b> .....	<b>1</b>
1.1 Scope .....	1
1.2 References .....	1
1.3 Test Equipment .....	1
1.4 Reference Ultrasonic Transducers .....	1
<b>Section 2 Terminology Related To Determining the Maximum Signal Level In The Test Set-Up</b> .....	<b>2</b>
2.1 Minimum Distance .....	2
2.2 Ultrasonic Transducer's Output .....	2
2.3 Test Distance .....	2
2.4 Maximum Voltage .....	2
2.5 Maximum Sound Pressure Level .....	2
<b>Section 3 Test Set-Up and Procedures</b> .....	<b>3</b>
3.1 General .....	3
3.2 Initial Test Set-Up .....	3
3.3 Hearing Aid Noise Test Procedure .....	4
<b>Tables</b>	
1-1 Example Reference Data for Ultrasonic Transducer .....	1
3-1 Maximum Sound Pressure Levels when 30 in from a Standard Ultrasonic Occupancy Sensor .....	3
<b>Figures</b>	
3-1 Schematic of Test Set-Up .....	3

## Foreword

This publication has been developed to address the potential for interference between hearing aids and ultrasonic lighting control devices (occupancy sensors). Some occupancy sensors may occasionally interfere with normal acoustic signal processing in some digital hearing instruments causing significant noise and distortion of the signal. This publication provides a basis to evaluate the possible interactions between ultrasonic lighting control devices and hearing aids utilizing a set of test procedures. This evaluation can be used as the basis for specifying performance criteria for both hearing aids and occupancy sensors to eliminate interference complaints.

The Lighting Controls Section of NEMA reviews this test procedure periodically for any revisions necessary to keep up to date with the latest technological advances. Proposed revisions or comments should be submitted to:

Senior Technical Director, Operations  
National Electrical Manufacturers Association  
1300 North 17<sup>th</sup> Street, Suite 900  
Rosslyn, VA 22209

Members of the NEMA Lighting Controls Section developed this Standards Publication. Section approval of this Standard does not necessarily imply that all section members voted for its approval or participated in its development. At the time it was approved, the Lighting Controls Section was composed of the following members:

Acuity Brands Lighting  
Eaton's Cooper Lighting  
GE Lighting  
Hubbell Building Automation  
Leviton Lighting & Energy Solutions  
Lutron Electronics Company, Inc.  
OSRAM SYLVANIA  
Philips Lighting Electronics North America  
RAB Lighting  
Schneider Electric  
Universal Lighting Technologies  
WattStopper

## Section 1 General

### 1.1 Scope

This Standards publication sets forth test procedures for use with a small acoustic chamber to evaluate potential interactions between hearing aids and ultrasonic lighting control devices, hereinafter referred to as occupancy sensors. The test procedures are designed to simulate and test occupancy sensors at three typical, specific frequencies (25 kHz, 32.7 kHz, and 40 kHz) and one type of hearing aid. However, if there are multiple hearing aids, the test procedures are repeated as many times as necessary.

### 1.2 References

This document is based on the report of David F. Henry and Barak Dar, *Effects of Ultrasonic Occupancy Sensors on Hearing Aids*, dated February 22, 2006.

### 1.3 Test Equipment

Test equipment to be used includes:

- a. 2CC acoustic coupler
- b. Acoustic test chamber
- c. Milli-voltmeter, or oscilloscope
- d. Ultrasonic occupancy sensor transducers
- e. Voltmeter
- f. Wave function generator

### 1.4 Reference Ultrasonic Transducers

Three sets of ultrasonic transducers (transmitter and receiver) are used to perform the compatibility testing. Each transducer set is tuned to one of the common ultrasonic occupancy sensors operating frequencies, i.e., 25, 32.768, or 40 kHz. The transmitter is calibrated to produce a known sound pressure level at the 12 in. (30 cm) specification distance and the receiver is calibrated to produce known electric output signal levels with specified input levels. An example of transducer data is shown in Table 1-1.

**Table 1-1  
Example Reference Data for Ultrasonic Transducers**

ID	Frequency	Transmitter Specification Voltage ( $V_{SPEC}$ )	Specification Distance – $D_{SPEC}$ (in. [cm])	Sound Level Meter - $SPL_{SPEC}$ (dB)	Receiver – $R_{SPEC}$ ( $V_{P-P}$ square wave)
25	25 kHz	20 VP-P [square wave]	12 in. (30 cm)	111	976
32	32.768 kHz			115	1200
40	40 kHz			117	1170